

References:

1. Ylönen J, Virtanen T, Rytkönen M, Mäntyjärvi R. Affinity purification of the major bovine allergen by a novel monoclonal antibody. *J Allergy Clin Immunol.* 1994 May;93(5):851-8.
2. Ylönen J, Virtanen T, Rytkönen M, Mäntyjärvi R. Quantification of a major bovine allergen by a two-site immunometric assay based on monoclonal antibodies. *Allergy.* 1994 Oct;49(9):707-12.
3. Rouvinen J, Rautiainen J, Virtanen T, Zeiler T, Kauppinen J, Taivainen A, Mäntyjärvi R. Probing the molecular basis of Allergy: Three-dimensional structure of the bovine lipocalin allergen Bos d 2. *The Journal of Biological Chemistry* 1999;274(4):2337-43.
4. Saarelainen SA, Kinnunen TT, Buhot C, Närvänen AT, Kauppinen AK, Rytkönen-Nissinen MA, Maillere B, Virtanen TI. Immunotherapeutic potential of the immunodominant T-cell epitope of lipocalin allergen Bos d 2 and its analogues. *Immunology.* 2008 Mar;123(3):358-66.
5. Filep S, Tsay A, Vailes LD, Gadermaier G, Ferreira F, Matsui EC et al. A multi-allergen standard for immunoassays: CREATE principles applied to eight purified allergens. *Allergy.* 18 Nov 2011; DOI: 10.1111/j. 1398-9995.2011.02750x.
6. Kauppinen A, Peräsaari J, Taivainen A, Kinnunen T, Saarelainen S, Rytkönen-Nissinen M, Jeal H, Jones M, Virtanen T. Association of HLA class II alleles with sensitization to cow dander Bos d 2, an important occupational allergen. *Immunobiology.* 2012 Jan; 217(1):8-12.




www.inbio.com

Indoor Biotechnologies, Inc.
1216 Harris Street,
Charlottesville,
VA 22903
United States

Tel: (434) 984-2304
Fax: (434) 984-2709
E-mail: mail@inbio.com

Indoor Biotechnologies Ltd.
Cardiff Medicentre
Heath Park
Cardiff CF14 4UJ
Wales

Tel: +44 (0) 2920 68 2115
Fax: +44 (0) 2920 68 2044
E-mail: info@indoorbiotech.co.uk

INDOOR and  are Registered Trademarks of INDOOR Biotechnologies, Ltd., Manchester U.K.




Bos d 2 ELISA kit (3D4/mAb1)

Product Code: EL-BD2

Lot Number: xxxxx

Sample Curve:



Content:

Vial 1 (red top) 100 μ L
Monoclonal antibody 3D4
Concentration: 1.33 mg/ml in PBS

Vial 2 (white top) 400 μ L
Bos d 2 Standard
Concentration: 500ng/ml Bos d 2

Vial 3 (brown) 100 μ L
Biotinylated monoclonal antibody mAb1
Dilute: 1:1000 for use

Storage: The ELISA kit should be stored at 4°C

For research and commercial use in vitro: not for human in vivo or therapeutic use.

Certificate of Analysis

Monoclonal Antibody: 3D4
Immunogen: Bovine dander allergen
Isotype: IgG1
Specificity: Binds to an epitope on bovine *Bos domesticus* dander allergen, Bos d 2.
Purification: Produced in tissue culture and purified by affinity chromatography using recombinant Protein G. Single heavy and light chain bands on SDS-PAGE.
Concentration: 1.33 mg/ml in phosphate buffered saline, pH 7.4. Based on A280 for IgG (1.42=1mg/ml), 0.22µm filtered, preservative free.
Lot Number: xxxxx

Monoclonal Antibody: mAb1
Immunogen: Bovine dander allergen
Isotype: IgG2b
Specificity: Binds to an epitope on bovine *Bos domesticus* dander allergen, Bos d 2.
Purification: Produced in tissue culture and purified by affinity chromatography using recombinant Protein G. Single heavy and light chain bands on SDS-PAGE.
Biotinylation: Biotinylated and titrated for use in ELISA at 1/1000 dilution. Prepared in 1% BSA/50% glycerol/PBS, pH 7.4, 0.22µm filtered, preservative free.
Lot Number: xxxxx

Allergen Standard: Bos d 2 Standard
Composition: Recombinant Bos d 2 allergen was prepared in 0.5% HSA/50% glycerol/PBS, pH 7.4.
Concentration: 500 ng/ml
Calibration: The concentration of the purified recombinant Bos d 2 was determined by amino acid analysis. There are no national or international reference standards for Bos d 2.
Lot Number: xxxxx

ELISA Protocol for Bos d 2.

1. Coat polystyrene microtiter plates (NUNC Maxisorp Cert. NUNC catalog # 439454) with 100µl mAb 3D4 at 10µl/10ml, i.e. 1/1000 dilution of stock, in 50mM carbonate-bicarbonate buffer, pH 9.6, incubate overnight at 4°C.
2. Wash wells 3x with PBS-0.05% Tween 20, pH 7.4 (PBS-T). Incubate for 1 hour at room temperature with 100µl/well of 1% BSA, PBS-T. Wash 3x with PBS-T.
3. Use doubling dilutions of the Bos d 2 Standard to make a control curve ranging from 50 - 0.1ng/ml: Pipette 20µl standard into 180µl 1% BSA, PBS-T into wells A1 and B1 on the ELISA plate. Mix well and transfer 100µl across the plate into 100µl 1% BSA, PBS-T diluent to make 10 serial doubling dilutions. Wells A11, B11 and A12, B12 should contain only 1% BSA, PBS-T as blanks.
4. Add 100µl of diluted allergen samples and incubate for 1 hour at room temperature. House dust extracts for Bos d 2 analysis are routinely diluted two-fold from 1/10-1/80. Other sample types, like air filter extracts and allergen extracts, may require different dilutions.
5. Wash wells 3x with PBS-T and add 100µl diluted biotinylated anti-Bos d 2 mAb mAb1. The antibody solution contains 50% glycerol and should be diluted 1/1000 in 1%BSA, PBS-T. Incubate for 1 hour at room temperature.

Wash wells 3x with PBS-T and add 100µl diluted Streptavidin-Peroxidase (Sigma S5512, 0.25mg reconstituted in 1ml distilled water). The reconstituted Streptavidin should be diluted 1/1000 in 1% BSA, PBS-T. Incubate for 30 minutes at room temperature.
7. Wash wells 3x and develop the assays by adding 100µl 1mM ABTS in 70mM citrate phosphate buffer, pH 4.2 and 1/1000 dilution of H₂O₂. Read the plate when the absorbance at 405nm reaches 2.0-2.4.

Notes:

The Bovine Allergen Standard is recommended for immunoassay calibration purposes only. Not recommended for in-vitro antibody measurements, T cell studies, immunization purposes, or other uses.

Buffer recipes, storage conditions and a list of frequently asked questions can be found under "Protocols" on our web site: www.inbio.com.

For research and commercial use in vitro: not for human in vivo or therapeutic use.